

MANUFACTURERS DECLARATION OF CONFORMITY

for

Product identification

Product : Biopotential measurement system
Brand : BioSemi
Model/Type : ActiveTwo
Version : ACT2 MOT16

Manufacturer : BioSemi B.V.
WG-Plein 129
1054 SC Amsterdam
Netherlands

Representative : Dr. A.C. Metting van Rijn
Function : Director

A sample of the product has been tested by :
D.A.R.E!! Consultancy
Vijzelmolenstraat 7
3447 GX Woerden
Netherlands

Standards used :

<i>Product Standard</i>	<i>Phenomenon</i>	<i>Results</i>
EN 61326 (2006)	Radiated emission	Pass
EN 61326 (1997) + A1 (1998) + A2 (2001)	ESD	Pass
EN 61326 (1997) + A1 (1998) + A2 (2001)	Radiated immunity	Pass
EN 60601-1 (1990) + A1 (1993) + A2 (1995) + A13 (1996)	Electrical safety	Pass

Test report no. : 10C00130RPT01, Mar 4, 2010 (EMC emission)
03C01115RPT01, Nov 19, 2003 (ESD & immunity)
04C00355RPT01, Apr 14, 2004 (Electrical Safety)

Means of conformity

The product is in conformity with Directives 89/336/EEC and 73/23/EEC based on the test results using harmonized standards in accordance with the directives.

Signature of representative of manufacturer :

Place : *Amsterdam*
Date : *24 June 2010*
Number : *2*

