



## DECLARATION OF CONFORMITY

EQUIPMENT TYPE/ENVIRONMENT: High-Speed Camera and Illuminators/ITE

TRADE NAME/MODEL NO.: EyeLink CL-Optilink

COUNTRY OF MANUFACTURE: Canada

APPLICATION OF COUNCIL DIRECTIVES:

- 89/336/EEC – The EMC Directive
- 2001/95/EC -- General Product Safety Directive
- 73/23/EEC -- Low Voltage Directive

STANDARDS TO WHICH CONFORMITY IS DECLARED:

- IEC/EN 60950-1:2005 (2nd edition)
- CAN/CSA-C22.2 No. 60950-00-CAN/CSA
- UL 60950 3<sup>rd</sup> edition
- CISPR 22:2003+A1:2004 / EN 55022:2003, Class A
- CISPR 24:1997 / EN 55024:1998+A1:2001+A2:2003
- IEC 60825-1:1993 + A1:1996 + A2:2001

MANUFACTURER'S NAME/ADDRESS:

SR Research Ltd.  
1570 Green Glade  
Mississauga, Ontario, Canada L5J 1B5

TEST LABORATIES:

UltraTech EMC Labs Inc (ITI (UK) accredited test facilities)  
3000 Bristol Circle  
Oakville, Ontario, Canada L6H 6G4

M.A. Lamonthe and Associates  
10 Oak Ridge Drive  
Georgetown Ontario, Canada L7G 5G6

QPS Evaluation Services Inc.  
81 Kelfield St. Unit 8  
Toronto, Canada M9W 5A3

CERTIFICATION BODY:

Intertek Testing services NA Inc.  
1829 32<sup>nd</sup> Avenue  
Lachine, Quebec, Canada H8T 3J1

As the Manufacturer, I declare that the equipment specified above conforms to the above named directives.

Dave Stampe, Vice President  
SR Research Ltd, Mississauga, Ontario Canada  
16 June 2009